

2 Test Summary

The customer requested EMC tests for a Pulse Rate Pedometer

| Test | Test Requirement | Test Method | Class / Severity | Result |
|-------------------------------------|--------------------|--|--|--------|
| Radiated Emission, 30MHz to 1GHz | EN 61000-6-3 :2007 | CISPR 16-2-3:2003+A2:2005 | Table 1 Column 3 of EN61000-6-3 | PASS |
| ESD | EN 61000-6-1: 2007 | EN 61000-4-2 :1995 + A1: 1998+A2:2001 | ±2, 4 kV Contact ±2, 4, 8 kV Air | PASS |
| Radiated Immunity | EN 61000-6-1: 2007 | EN 61000-4-3 :2006 | (80MHz to 1GHz) 3V/m, 80%, 1kHz Amp. Mod. (1.4GHz to 2GHz) 3V/m, 80%, 1kHz Amp. Mod. (2GHz to 2.7GHz) 1V/m, 80%, 1kHz Amp. Mod. | PASS |

This document is issued by the Company under its General Conditions of Service printed overleaf or available on request and accessible at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law." Unless otherwise stated the results shown in this test report refer only to the sample(s) tested. This document cannot be reproduced except in full, without prior approval of the Company.